



Contribution ID: 57

Type: **Poster Presentation**

## Performance of additively manufactured ZnMg for biomedical implant applications

The newest research for alternatives to autologous bone grafts when treating bone defects leads to the development of bioresorbable materials. In this work, bioresorbable zinc-magnesium alloys are 3D-printed by Powder Bed Fusion of Metals using a Laser Beam (PBF-LB/M). The goal is to manufacture load-bearing implants (e. g. patient-specific scaffolds or cages) with suitable degradation properties and non-toxic byproducts, avoiding harm to the surrounding tissue. The mechanical properties of bulky and scaffold-like structures as well as their biocompatibility according to ISO 10993-5 are investigated.

### Speaker Country

Germany

**Primary authors:** FISCHER, Florian (DAP (Digital Additive Production) - RWTH Aachen University); VOSHAGE, Maximilian (RWTH Aachen University - DAP)

**Co-authors:** Dr JAUER, Lucas (DAP (Digital Additive Production) - RWTH Aachen University); Mr ZHAO, Qun (Experimental Orthopaedics and Trauma Surgery, Department of Orthopaedics, Trauma and Reconstructive Surgery, University Hospital RWTH Aachen); Dr KOPP, Alexander (meotec GmbH); Dr GREVEN, Johannes (Department of Thoracic and Cardiovascular Surgery, Medical Faculty, RWTH Aachen, University Hospital Aachen); Prof. SCHLEIFENBAUM, Johannes Henrich (RWTH Aachen University, Chair for Digital Additive Production DAP, Germany)

**Presenter:** FISCHER, Florian (DAP (Digital Additive Production) - RWTH Aachen University)

**Track Classification:** Tools, Space and Aircraft, Automotive, Medical and others